

Localization – Membrane

Host Species – Rabbit

Ig Class – IgG

Intended Use

This antibody is designed for the specific localization of CD43 in formalin-fixed, paraffin-embedded (FFPE) tissue sections.

Storage & Handling

Store RTU Vial at 2-8°C. Fresh dilutions for concentrated antibodies, if required, should be prepared prior to use and are stable for up to one day at room temperature (20-26°C).

Working Principle

IHC is a two-step process wherein the primary antibody binds to the antigen of interest and that binding is detected by a chromogen. The primary antibody may be used in IHC using manual techniques or any Automated Staining System. Positive and negative controls should always be run simultaneously with all patient specimens.

Product Description

CD43, also referred to as leukosialin or sialophorin, is a heavily glycosylated, mucin-like transmembrane protein distinguished by its highly anionic extracellular domain. It is broadly expressed across hematopoietic lineages, with the exception of quiescent B lymphocytes. CD43 exhibits both pro-adhesive and anti-adhesive properties, forming a steric and electrostatic barrier that can inhibit cell-cell interactions while also facilitating leukocyte aggregation under specific physiological conditions. It plays a role in the activation and functional regulation of T cells, B cells, natural killer (NK) cells, and monocytes. Upon cellular activation, CD43 expression is downregulated through proteolytic cleavage. CD43 engages in trans-interactions with its counter-receptor, CD169 (SIGLEC-1), which is expressed on macrophages. Intracellularly, CD43 associates with the ERM (ezrin-radixin-moesin) family of actin-binding proteins, contributing to cytoskeletal rearrangement, T-cell polarization, and leukocyte trafficking. It modulates integrin activity, particularly LFA-1, enhancing its adhesiveness while antagonizing other LFA-1 regulatory pathways. CD43 signaling can induce p53 tumor suppressor activity; however, in the absence of functional p53 or ARF, it may facilitate tumor progression. Dysregulation of CD43 has been implicated in immunodeficiencies and malignancies, including Wiskott

Material Supplied

CD43 antibody is affinity purified and diluted in PBS, pH 7.4, containing 1% BSA and 0.09% sodium azide.

Material required But Not Supplied

- Xylene
- DI Water
- Control Tissues
- Isopropyl alcohol
- Antigen retrieval buffers
- Hematoxylin
- Positive charged slides
- Blocking Reagents
- Mounting media
- Wash Buffer
- Detection System
- Cover glass

Working Reagent Procedure

- Ready-to-Use antibodies have been optimized for use with the recommended protocols and should not require further dilution.
- Concentrated antibodies must be diluted in accordance with the recommended protocol.

Recommended Protocol

Refer the following table for the details on specific recommended protocol for this antibody.

Control Tissue	Lymph Node, Tonsil	Antibody Incubation Time	30-60 Minutes at RT
Dilution factor	1:20-50 (Antibody Diluent: DH144)	Retrieval Pre-treatment	Tris-EDTA based HIER (AR9 Buffer: DH020)

Precautions

This product should be used by qualified and trained professional users only.

Avoid microbial contamination of reagents to minimize non-specific staining. Never pipette reagents by mouth. Avoid contact of reagents and specimens with skin. If reagents or specimens come into contact with sensitive area, wash with sufficient amounts of water. Dispose of the unused reagents. This kit contain sodium azide at concentrations of less than 0.1%. Sodium azide is not classified as a hazardous chemical at these concentrations, but proper handling protocols should be observed. For more information on product hazards, precautions and waste disposal, *Material Safety Data Sheets* are available upon request.







Limitations

Improper tissue handling and processing prior to immunostaining can lead to inconsistent results. Variations in embedding and fixation or the nature of the tissue may lead to variations in results. Endogenous peroxidase activity or pseudo peroxidase activity in erythrocytes and tissue biotin may result in non-specific staining based on the detection system employed. Tissues containing Hepatitis B Surface Antigen (HBsAg) may give false positive with horseradish peroxidase systems. Improper counterstaining and mounting may compromise the interpretation of results. Interpretation of the staining result is solely the responsibility of the user. Experimental results should be confirmed by a medically-established diagnostic product or procedure. Evaluation must be performed by a qualified pathologist.

Troubleshooting

For Technical Support contact us at +91 - 7506501122 or info@dygnova.com or your local distributor to report unusual staining.

Doc No: DH/DS/CD373Rev.00

	Manufacturer Details		Use by Date	LOT	Lot/Batch Number
	Manufacturing Date		Consult Instructions for Use	REF	Catalogue Number
	Temperature Limits		Sufficient for 'n' assays / tests	IVD	In-vitro Diagnostic Medical Device